

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2587

Name: RITTER2  
Sterilizer ID: V2313228  
06/29/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
2:00 200.3 F 4.2  
4:00 208.7 F 10.0  
6:00 222.4 F 6.7  
8:00 234.5 F 11.8  
10:00 236.2 F 9.9  
12:00 236.7 F 8.8  
14:00 250.1 F 15.7  
16:00 262.1 F 22.6  
18:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.6 F 29.6  
2:00 271.6 F 28.8  
3:00 272.3 F 29.7  
4:00 272.4 F 29.3  
  
Min 271.3 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 213.6 F 0.0  
10:00 224.0 F 0.0  
15:00 224.0 F 0.0  
20:00 225.1 F 0.0  
25:00 223.0 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 18:23  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:43  
CYCLE COMPLETE

06/29/2023 12:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_