

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1438

Name: RITTER 1  
Sterilizer ID: V2178198  
12/09/2021 07:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 81.8 F 0.0  
4:00 122.1 F 0.0  
6:00 172.3 F 0.1  
8:00 178.1 F 0.1  
10:00 195.8 F 0.3  
12:00 210.5 F 0.9  
14:00 221.8 F 4.2  
16:00 228.8 F 6.6  
18:00 236.7 F 9.9  
20:00 240.8 F 11.1  
22:00 242.6 F 11.5  
24:00 238.4 F 9.5  
26:00 237.6 F 8.7  
28:00 250.6 F 15.1  
30:00 262.3 F 21.8  
32:00 272.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.3 F 28.1  
2:00 272.8 F 28.6  
3:00 271.6 F 27.8  
4:00 272.5 F 28.3  
  
Min 271.5 F 27.6  
Max 272.9 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 216.9 F 0.0  
10:00 200.5 F 0.0  
15:00 205.6 F 0.0  
20:00 191.8 F 0.0  
25:00 188.7 F 0.0  
30:00 187.0 F 0.0  
  
Drying Complete

FILLING: 1:02  
HEATING: 32:19  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:10:26  
CYCLE COMPLETE

12/09/2021 09:03 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02