
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006987
16:35 26/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:27
136.0 C 220kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
136.4 C 222kPa 5:57
136.2 C 224kPa 6:27
136.4 C 221kPa 6:57
136.4 C 225kPa 7:27
136.3 C 223kPa 7:57
136.3 C 223kPa 7:57
VENTING 7:58
AIR DRYING 8:49
CYCLE COMPLETE 9:02

Digital Signature #
650004463670ACED
