

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5214

Name: RITTER 1  
Sterilizer ID: V2178198  
12/30/2024 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.9 F 0.0  
2:00 71.9 F 0.0  
4:00 81.8 F 0.0  
6:00 119.2 F 0.0  
8:00 155.8 F 0.1  
10:00 182.0 F 0.2  
12:00 199.0 F 0.4  
14:00 214.2 F 2.7  
16:00 223.3 F 5.7  
18:00 230.8 F 8.6  
20:00 239.7 F 12.3  
22:00 235.5 F 9.0  
24:00 240.1 F 10.4  
26:00 245.4 F 13.1  
28:00 243.0 F 11.8  
30:00 256.0 F 18.5  
32:00 267.1 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.7 F 29.0  
2:00 271.8 F 28.2  
3:00 272.3 F 28.9  
4:00 272.5 F 28.7  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 198.6 F 0.0  
10:00 191.3 F 0.0  
15:00 200.9 F 0.0  
20:00 202.4 F 0.0  
25:00 208.2 F 0.0  
30:00 215.7 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:32  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:11:40  
CYCLE COMPLETE

12/30/2024 12:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_