
Midmark M11 - v1.0.5
Total Cycles: 1251

Name: RITTER2
Sterilizer ID: V2313228
07/04/2022 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 187.6 F 3.5
4:00 188.2 F 8.0
6:00 214.9 F 13.2
8:00 225.6 F 8.5
10:00 239.0 F 11.7
12:00 239.0 F 11.3
14:00 237.2 F 9.8
16:00 250.7 F 16.4
18:00 262.2 F 23.1
20:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.8
1:00 272.8 F 30.5
2:00 272.4 F 29.8
3:00 272.6 F 30.6
4:00 272.9 F 30.1

Min 271.9 F 29.7
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 211.3 F 0.0
10:00 226.1 F 0.0
15:00 230.7 F 0.0
20:00 227.8 F 0.0
25:00 225.8 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

07/04/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____