
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001437
17:50 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:42
136.5 C 225kPa 3:42
136.4 C 226kPa 4:12
136.3 C 225kPa 4:42
136.4 C 226kPa 5:12
136.2 C 225kPa 5:42
136.4 C 226kPa 6:12
136.3 C 225kPa 6:42
136.3 C 225kPa 7:12
136.3 C 225kPa 7:12
VENTING 7:12
AIR DRYING 8:06
CYCLE COMPLETE 12:00

Digital Signature #
DC09693E75CF7C9D
