

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000692  
8:00 16/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:07  
PRESSURIZING 2:13  
STERILIZING 3:18  
136.0 C 228kPa 3:18  
136.1 C 230kPa 3:48  
136.0 C 230kPa 4:18  
136.1 C 230kPa 4:48  
136.2 C 230kPa 5:18  
136.4 C 232kPa 5:48  
136.4 C 231kPa 6:18  
136.3 C 231kPa 6:48  
136.3 C 231kPa 6:48  
VENTING 6:48  
AIR DRYING 7:39  
CYCLE COMPLETE 7:41  
-----

Digital Signature #  
58960F708004BC3F  
-----