
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001448
8:55 09/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:20
PRESSURIZING 7:09
STERILIZING 8:55
136.4 C 221kPa 8:55
136.2 C 221kPa 9:25
136.4 C 220kPa 9:55
136.4 C 221kPa 10:25
136.5 C 222kPa 10:55
136.4 C 221kPa 11:25
135.9 C 218kPa 11:55
136.4 C 223kPa 12:25
VENTING 12:25
AIR DRYING 13:27
CYCLE COMPLETE 43:27

Digital Signature #
E922857171A4AA11
