
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001571
15:47 07/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:23
STERILIZING 4:28
136.0 C 221kPa 4:28
136.2 C 224kPa 4:58
136.4 C 226kPa 5:28
136.3 C 225kPa 5:58
136.4 C 226kPa 6:28
136.3 C 225kPa 6:58
136.3 C 225kPa 7:28
136.4 C 223kPa 7:58
136.4 C 223kPa 7:58
VENTING 7:58
AIR DRYING 8:53
CYCLE COMPLETE 9:03

Digital Signature #
E4C8FB04C7851C80
