
Midmark M11 - v1.0.0
Total Cycles: 2830

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.7 F 0.0
2:00 80.2 F 0.0
4:00 93.0 F 0.1
6:00 122.1 F 1.7
8:00 152.6 F 4.6
10:00 173.0 F 8.0
12:00 195.0 F 11.6
14:00 208.4 F 3.3
16:00 217.4 F 7.1
18:00 228.8 F 10.8
20:00 231.0 F 7.3
22:00 239.9 F 11.4
24:00 236.6 F 8.8
26:00 249.6 F 14.9
28:00 260.4 F 21.0
30:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.6 F 28.3
3:00 271.8 F 28.8
4:00 272.2 F 28.8

Min 271.2 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 206.8 F 0.0
10:00 215.8 F 0.0
15:00 216.9 F 0.0
20:00 219.6 F 0.0
25:00 223.2 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:51
CYCLE COMPLETE

01/31/2023 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____