

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001181  
13:27 21/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:56  
STERILIZING 4:04  
136.0 C 220kPa 4:04  
136.0 C 219kPa 4:34  
136.2 C 222kPa 5:04  
136.3 C 224kPa 5:34  
136.4 C 225kPa 6:04  
136.4 C 225kPa 6:34  
136.5 C 225kPa 7:04  
136.2 C 221kPa 7:34  
136.2 C 221kPa 7:34  
VENTING 7:34  
AIR DRYING 8:24  
CYCLE COMPLETE 9:16

-----  
Digital Signature #  
ADBA178B064F81CF  
-----