

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006558  
10:28 06/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:36  
PRESSURIZING 3:36  
STERILIZING 4:35  
136.4 C 220kPa 4:35  
136.3 C 223kPa 5:05  
136.3 C 224kPa 5:35  
136.2 C 222kPa 6:05  
136.4 C 222kPa 6:35  
136.4 C 223kPa 7:05  
136.3 C 224kPa 7:35  
135.8 C 220kPa 8:05  
135.8 C 220kPa 8:05  
VENTING 8:05  
AIR DRYING 8:56  
CYCLE COMPLETE 9:12

-----  
Digital Signature #  
2DF54F0BE6A89741  
-----