
Midmark M11 - v1.0.0
Total Cycles: 2257

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.0 F 0.0
2:00 204.0 F 0.5
4:00 211.7 F 3.0
6:00 223.5 F 8.1
8:00 236.0 F 12.9
10:00 237.0 F 9.7
12:00 239.5 F 9.2
14:00 245.0 F 12.1
16:00 246.2 F 13.1
18:00 258.5 F 19.7
20:00 269.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.7
2:00 271.6 F 27.9
3:00 272.5 F 28.9
4:00 272.1 F 28.4

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 235.5 F 0.0
10:00 227.2 F 0.0
15:00 230.9 F 0.0
20:00 226.5 F 0.0
25:00 230.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

08/29/2022 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____