
Midmark M11 - v1.0.0
Total Cycles: 640

Name: MIDMARK
Sterilizer ID: V2178198
01/28/2021 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 187.2 F 1.8
4:00 196.6 F 4.9
6:00 214.0 F 9.2
8:00 226.5 F 9.7
10:00 232.4 F 10.0
12:00 235.6 F 8.0
14:00 241.4 F 12.3
16:00 240.2 F 10.2
18:00 251.7 F 16.4
20:00 262.4 F 22.6
22:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.2
1:00 272.1 F 29.3
2:00 272.0 F 29.2
3:00 272.6 F 29.2
4:00 272.4 F 29.3

Min 271.7 F 29.1
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 229.4 F 0.0
10:00 235.2 F 0.0
15:00 230.3 F 0.0
20:00 231.0 F 0.0
25:00 230.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

01/28/2021 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____