
Midmark M11 - v1.0.0
Total Cycles: 3852

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 186.5 F 4.0
4:00 205.8 F 9.3
6:00 219.3 F 7.8
8:00 230.8 F 9.4
10:00 236.7 F 9.5
12:00 244.3 F 12.9
14:00 243.7 F 12.0
16:00 256.6 F 18.8
18:00 268.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.6
2:00 271.8 F 28.4
3:00 272.5 F 28.6
4:00 271.5 F 28.2

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 210.2 F 0.0
10:00 224.1 F 0.0
15:00 226.3 F 0.0
20:00 225.3 F 0.0
25:00 222.3 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

10/02/2023 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____