
Midmark M11 - v1.0.0
Total Cycles: 4953

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 150.5 F 0.7
4:00 154.1 F 1.4
6:00 184.5 F 4.8
8:00 205.4 F 8.6
10:00 220.0 F 12.5
12:00 223.6 F 6.7
14:00 231.8 F 11.1
16:00 232.2 F 6.9
18:00 241.3 F 12.5
20:00 239.9 F 10.3
22:00 251.4 F 16.0
24:00 261.4 F 21.7
26:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 272.2 F 28.9
3:00 271.5 F 28.8
4:00 272.5 F 29.4

Min 271.4 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 207.5 F 0.0
10:00 229.2 F 0.0
15:00 234.3 F 0.0
20:00 229.8 F 0.0
25:00 228.9 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:44
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:06:23
CYCLE COMPLETE

10/02/2024 12:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____