
Midmark M11 - v1.0.5
Total Cycles: 3614

Name: RITTER2
Sterilizer ID: V2313228
05/14/2024 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 138.9 F 1.5
4:00 147.5 F 3.9
6:00 172.9 F 7.2
8:00 205.9 F 11.2
10:00 220.0 F 4.3
12:00 227.7 F 8.3
14:00 239.3 F 12.6
16:00 235.0 F 9.1
18:00 239.9 F 11.0
20:00 244.5 F 12.9
22:00 255.7 F 18.7
24:00 265.4 F 24.8
26:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.2 F 29.4
2:00 273.0 F 29.8
3:00 271.8 F 29.8
4:00 271.7 F 29.3

Min 271.2 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 233.8 F 0.0
10:00 229.2 F 0.0
15:00 196.4 F 0.0
20:00 181.9 F 0.0
25:00 183.8 F 0.0
30:00 181.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:37
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:09
CYCLE COMPLETE

05/14/2024 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____