
Midmark M11 - v1.0.0
Total Cycles: 2007

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2022 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.5 F 0.0
2:00 152.2 F 2.2
4:00 162.0 F 5.3
6:00 211.7 F 10.2
8:00 223.0 F 4.1
10:00 225.4 F 7.9
12:00 237.0 F 12.9
14:00 236.5 F 9.5
16:00 239.4 F 9.1
18:00 246.9 F 13.8
20:00 259.1 F 20.3
22:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.3
1:00 272.4 F 29.0
2:00 272.5 F 28.9
3:00 272.7 F 29.4
4:00 271.9 F 28.7

Min 270.9 F 28.1
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 221.1 F 0.0
10:00 231.4 F 0.0
15:00 230.5 F 0.0
20:00 228.3 F 0.0
25:00 230.7 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

06/08/2022 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____