
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002735
8:55 29/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:16
STERILIZING 4:27
136.4 C 223kPa 4:27
136.1 C 222kPa 4:57
136.3 C 223kPa 5:27
136.1 C 220kPa 5:57
136.3 C 220kPa 6:27
136.4 C 222kPa 6:57
136.2 C 221kPa 7:27
136.3 C 221kPa 7:57
136.3 C 221kPa 7:57
VENTING 7:57
AIR DRYING 8:46
CYCLE COMPLETE 13:11

Digital Signature #
C7AA4ADF0B25CDFF
