

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3082

Name: RITTER 1  
Sterilizer ID: V2178198  
04/10/2023 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
2:00 197.3 F 3.8  
4:00 212.5 F 9.1  
6:00 228.2 F 7.4  
8:00 234.5 F 10.5  
10:00 235.9 F 7.6  
12:00 243.4 F 11.7  
14:00 245.8 F 13.0  
16:00 258.5 F 19.9  
18:00 269.9 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.7  
2:00 272.2 F 28.8  
3:00 272.2 F 28.5  
4:00 272.1 F 28.8  
  
Min 271.5 F 28.1  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 233.8 F 0.0  
10:00 231.1 F 0.0  
15:00 227.0 F 0.0  
20:00 227.2 F 0.0  
25:00 230.9 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 2:14  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

04/10/2023 01:14 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_