

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1283

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2021 08:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.7 F 0.0  
2:00 166.7 F 1.6  
4:00 176.6 F 4.9  
6:00 203.4 F 9.7  
8:00 217.4 F 5.6  
10:00 231.3 F 9.8  
12:00 236.4 F 8.7  
14:00 242.7 F 12.8  
16:00 241.8 F 11.3  
18:00 255.0 F 18.1  
20:00 266.5 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.3 F 29.1  
2:00 272.2 F 28.9  
3:00 271.4 F 28.4  
4:00 273.0 F 29.4  
  
Min 271.2 F 28.3  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 228.1 F 0.0  
10:00 221.1 F 0.0  
15:00 234.0 F 0.0  
20:00 225.3 F 0.0  
25:00 220.0 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 21:24  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:21  
CYCLE COMPLETE

10/21/2021 09:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02