
Midmark M11 - v1.0.0
Total Cycles: 5259

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 165.6 F 3.1
4:00 182.5 F 7.0
6:00 199.9 F 11.2
8:00 215.0 F 3.6
10:00 226.2 F 8.2
12:00 237.1 F 12.7
14:00 235.8 F 9.1
16:00 242.0 F 10.5
18:00 246.2 F 13.0
20:00 257.6 F 19.2
22:00 267.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.7 F 28.6
2:00 272.4 F 28.7
3:00 271.7 F 28.1
4:00 271.5 F 28.4

Min 271.2 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 214.2 F 0.0
10:00 202.9 F 0.0
15:00 200.9 F 0.0
20:00 197.8 F 0.0
25:00 209.1 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

01/22/2025 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____