
Midmark M11 - v1.0.0
Total Cycles: 4750

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2024 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 175.2 F 3.5
4:00 190.8 F 7.3
6:00 218.6 F 12.0
8:00 221.5 F 5.4
10:00 232.4 F 9.9
12:00 236.4 F 8.7
14:00 241.2 F 12.0
16:00 239.7 F 10.3
18:00 251.9 F 16.3
20:00 262.8 F 22.6
22:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.9
2:00 271.2 F 28.2
3:00 272.5 F 29.0
4:00 271.9 F 28.6

Min 271.1 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 190.4 F 0.0
10:00 216.8 F 0.0
15:00 223.1 F 0.0
20:00 219.7 F 0.0
25:00 214.3 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:01:35
CYCLE COMPLETE

07/19/2024 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____