
Midmark M11 - v1.0.0
Total Cycles: 4586

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2024 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 82.2 F 0.0
2:00 83.2 F 0.5
4:00 101.0 F 2.0
6:00 129.2 F 4.1
8:00 156.2 F 6.5
10:00 175.9 F 9.3
12:00 187.0 F 12.4
14:00 204.0 F 4.0
16:00 214.7 F 7.4
18:00 223.5 F 11.3
20:00 227.3 F 6.5
22:00 238.6 F 11.6
24:00 236.3 F 9.3
26:00 248.4 F 15.0
28:00 258.9 F 20.7
30:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.6
2:00 272.1 F 29.2
3:00 271.1 F 28.8
4:00 271.7 F 29.4

Min 271.1 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 226.2 F 0.0
10:00 224.2 F 0.0
15:00 224.4 F 0.0
20:00 233.7 F 0.0
25:00 228.6 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

05/27/2024 04:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____