
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001495
14:17 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:49
STERILIZING 3:54
136.5 C 223kPa 3:54
136.1 C 223kPa 4:24
136.4 C 224kPa 4:54
136.4 C 225kPa 5:24
136.3 C 223kPa 5:54
136.4 C 223kPa 6:24
136.2 C 223kPa 6:54
136.4 C 225kPa 7:24
136.4 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 38:14

Digital Signature #
316284D1168D1197
