
Midmark M11 - v1.0.5
Total Cycles: 1219

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 184.2 F 3.7
4:00 188.7 F 8.2
6:00 210.8 F 13.3
8:00 225.3 F 8.4
10:00 235.7 F 12.0
12:00 240.4 F 11.8
14:00 239.7 F 10.6
16:00 252.9 F 17.3
18:00 264.3 F 24.2
20:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.9 F 29.8

Min 271.2 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 216.3 F 0.0
10:00 223.0 F 0.0
15:00 223.8 F 0.0
20:00 220.9 F 0.0
25:00 219.2 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

06/24/2022 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____