

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 004774  
14:27 28/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:31  
STERILIZING 4:36  
136.4 C 221kPa 4:36  
136.4 C 226kPa 5:06  
136.4 C 224kPa 5:36  
136.4 C 222kPa 6:06  
136.3 C 224kPa 6:36  
136.3 C 223kPa 7:06  
136.4 C 223kPa 7:36  
136.4 C 224kPa 8:06  
136.4 C 224kPa 8:06  
VENTING 8:06  
AIR DRYING 9:00  
CYCLE COMPLETE 9:02

-----  
Digital Signature #  
11FE476295A09FA2  
-----