
Midmark M11 - v1.0.5
Total Cycles: 409

Name: RITTER2
Sterilizer ID: V2313228
09/30/2021 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.0 F 0.0
2:00 98.9 F 0.6
4:00 123.0 F 2.4
6:00 173.2 F 5.5
8:00 193.8 F 9.5
10:00 206.8 F 9.4
12:00 222.2 F 6.8
14:00 233.1 F 11.5
16:00 230.0 F 7.8
18:00 243.0 F 13.3
20:00 241.7 F 11.8
22:00 253.4 F 17.8
24:00 263.9 F 24.1
26:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.3 F 29.5
2:00 272.6 F 29.9
3:00 271.8 F 29.7
4:00 272.0 F 29.4

Min 271.5 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 223.7 F 0.0
10:00 214.5 F 0.0
15:00 220.5 F 0.0
20:00 212.0 F 0.0
25:00 211.5 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 26:06
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

09/30/2021 10:03 AM
ACCEPTED: X
REJECTED:
Operator: U02