
Midmark M11 - v1.0.0
Total Cycles: 4090

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2023 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 142.5 F 2.2
4:00 160.6 F 5.6
6:00 188.0 F 9.6
8:00 214.4 F 5.5
10:00 220.4 F 7.6
12:00 232.9 F 12.3
14:00 234.4 F 8.6
16:00 242.1 F 11.4
18:00 244.8 F 12.5
20:00 256.9 F 18.7
22:00 267.3 F 25.3
24:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.7 F 28.8
2:00 272.7 F 29.2
3:00 271.6 F 28.5
4:00 272.5 F 29.0

Min 271.5 F 28.4
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 236.1 F 0.0
10:00 196.7 F 0.0
15:00 185.2 F 0.0
20:00 181.4 F 0.0
25:00 179.3 F 0.0
30:00 179.5 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

12/06/2023 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____