
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002075
7:18 20/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 3:01
136.0 C 220kPa 3:01
136.3 C 223kPa 3:31
136.2 C 222kPa 4:01
136.4 C 222kPa 4:31
136.4 C 222kPa 5:01
136.4 C 223kPa 5:31
136.3 C 222kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 6:31
VENTING 6:32
AIR DRYING 7:26
CYCLE COMPLETE 7:37

Digital Signature #
FFECDEEA5B8B3238
