
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003555
10:32 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:37
STERILIZING 3:29
136.0 C 220kPa 3:29
136.3 C 225kPa 3:59
136.2 C 221kPa 4:29
136.2 C 223kPa 4:59
136.3 C 223kPa 5:29
136.4 C 223kPa 5:59
136.3 C 224kPa 6:29
136.2 C 224kPa 6:59
136.2 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:48
CYCLE COMPLETE 9:01

Digital Signature #
8D7C33F89C53CDA3
