
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001633
13:51 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:39
STERILIZING 3:41
136.4 C 227kPa 3:41
135.9 C 222kPa 4:11
136.4 C 226kPa 4:41
136.4 C 226kPa 5:11
136.4 C 224kPa 5:41
136.1 C 222kPa 6:11
136.1 C 223kPa 6:41
136.4 C 223kPa 7:11
136.4 C 223kPa 7:11
VENTING 7:11
AIR DRYING 8:03
CYCLE COMPLETE 8:12

Digital Signature #
11AE37B20C583B84
