
Midmark M11 - v1.0.5
Total Cycles: 4094

Name: RITTER2
Sterilizer ID: V2313228
11/06/2024 04:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 150.8 F 2.1
4:00 162.5 F 5.7
6:00 187.2 F 10.2
8:00 211.2 F 5.3
10:00 223.9 F 9.5
12:00 232.0 F 9.5
14:00 239.1 F 11.8
16:00 238.6 F 10.3
18:00 252.1 F 16.9
20:00 263.7 F 23.9
22:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.9
2:00 271.8 F 29.1
3:00 272.2 F 29.9
4:00 272.4 F 29.6

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 223.3 F 0.0
10:00 203.5 F 0.0
15:00 198.6 F 0.0
20:00 195.5 F 0.0
25:00 193.6 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

11/06/2024 05:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____