
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000582
7:08 06/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:02
STERILIZING 4:29
136.4 C 220kPa 4:29
136.2 C 221kPa 4:59
136.0 C 217kPa 5:29
136.4 C 223kPa 5:59
136.4 C 222kPa 6:29
136.4 C 222kPa 6:59
136.3 C 223kPa 7:29
136.3 C 220kPa 7:59
136.3 C 220kPa 7:59
VENTING 7:59
AIR DRYING 8:54
CYCLE COMPLETE 9:52

Digital Signature #
C0530AEA8F817F70
