
Midmark M11 - v1.0.5
Total Cycles: 2286

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.8 F 0.0
2:00 153.0 F 2.8
4:00 170.3 F 6.5
6:00 192.8 F 11.4
8:00 215.2 F 5.7
10:00 231.4 F 11.0
12:00 231.0 F 7.9
14:00 240.8 F 11.6
16:00 244.9 F 13.3
18:00 257.5 F 20.1
20:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.3 F 30.0
2:00 272.0 F 29.3
3:00 271.9 F 29.8
4:00 272.4 F 29.6

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 217.5 F 0.0
10:00 221.5 F 0.0
15:00 207.9 F 0.0
20:00 211.3 F 0.0
25:00 232.3 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

04/21/2023 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____