

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4015

Name: RITTER2  
Sterilizer ID: V2313228  
10/09/2024 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.2 F 0.0  
2:00 121.9 F 0.0  
4:00 125.6 F 0.3  
6:00 157.0 F 3.4  
8:00 173.5 F 7.1  
10:00 193.2 F 10.9  
12:00 217.3 F 5.7  
14:00 220.2 F 9.6  
16:00 233.3 F 10.1  
18:00 236.9 F 11.7  
20:00 238.0 F 10.1  
22:00 250.3 F 16.1  
24:00 260.8 F 22.2  
26:00 270.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.8  
1:00 271.9 F 29.7  
2:00 272.1 F 29.7  
3:00 272.1 F 29.8  
4:00 272.1 F 30.0

Min 271.7 F 29.7  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 215.1 F 0.0  
10:00 204.3 F 0.0  
15:00 188.0 F 0.0  
20:00 185.1 F 0.0  
25:00 185.0 F 0.0  
30:00 186.6 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:31  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:06:14  
CYCLE COMPLETE

10/09/2024 05:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_