
Midmark M11 - v1.0.5
Total Cycles: 2943

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.4 F 0.0
2:00 123.1 F 0.0
4:00 142.2 F 0.0
6:00 176.5 F 2.8
8:00 196.7 F 6.6
10:00 216.5 F 10.9
12:00 222.4 F 6.3
14:00 234.6 F 11.6
16:00 234.5 F 9.0
18:00 239.3 F 9.9
20:00 247.8 F 14.4
22:00 259.9 F 21.2
24:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.4 F 29.5
2:00 271.4 F 28.7
3:00 272.5 F 29.4
4:00 271.4 F 28.7

Min 271.3 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 211.0 F 0.0
10:00 222.1 F 0.0
15:00 212.8 F 0.0
20:00 213.2 F 0.0
25:00 223.3 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

09/28/2023 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____