
Midmark M11 - v1.0.5
Total Cycles: 3692

Name: RITTER2
Sterilizer ID: V2313228
06/12/2024 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.0 F 0.0
2:00 168.4 F 3.0
4:00 181.6 F 6.9
6:00 211.0 F 11.4
8:00 218.5 F 4.8
10:00 232.8 F 9.1
12:00 236.1 F 11.1
14:00 241.2 F 11.7
16:00 239.8 F 10.6
18:00 252.6 F 17.0
20:00 263.7 F 23.6
22:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.2
2:00 271.5 F 29.1
3:00 272.5 F 29.2
4:00 271.5 F 29.0

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 224.3 F 0.0
10:00 195.6 F 0.0
15:00 193.0 F 0.0
20:00 189.8 F 0.0
25:00 187.5 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

06/12/2024 04:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____