

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008183  
10:23 19/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:02  
STERILIZING 2:56  
136.0 C 220kPa 2:56  
136.2 C 220kPa 3:26  
136.4 C 224kPa 3:56  
136.1 C 217kPa 4:26  
136.4 C 224kPa 4:56  
136.3 C 222kPa 5:26  
136.4 C 224kPa 5:56  
136.4 C 224kPa 6:26  
136.4 C 224kPa 6:26  
VENTING 6:26  
AIR DRYING 7:16  
CYCLE COMPLETE 37:16

-----  
Digital Signature #  
5D2B4B009FC31119  
-----