
Midmark M11 - v1.0.0
Total Cycles: 1018

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 184.1 F 3.1
4:00 193.3 F 7.2
6:00 211.2 F 12.4
8:00 224.8 F 7.3
10:00 237.8 F 12.6
12:00 236.8 F 10.0
14:00 236.3 F 8.6
16:00 249.5 F 15.0
18:00 261.5 F 21.7
20:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.2
3:00 272.4 F 29.1
4:00 272.0 F 28.5

Min 271.3 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 220.6 F 0.0
10:00 233.6 F 0.0
15:00 231.3 F 0.0
20:00 227.4 F 0.0
25:00 227.3 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

07/15/2021 12:05 PM
ACCEPTED: X
REJECTED:
Operator: U02