
Midmark M11 - v1.0.5
Total Cycles: 3502

Name: RITTER2
Sterilizer ID: V2313228
04/05/2024 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.7 F 0.0
2:00 147.0 F 2.2
4:00 170.5 F 5.6
6:00 195.6 F 9.7
8:00 213.9 F 8.3
10:00 219.4 F 8.4
12:00 232.7 F 13.4
14:00 235.6 F 10.7
16:00 235.9 F 9.2
18:00 249.7 F 15.8
20:00 261.3 F 22.4
22:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.8 F 30.2
2:00 271.6 F 29.5
3:00 271.8 F 30.0
4:00 272.5 F 30.1

Min 271.2 F 29.3
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 231.9 F 0.0
10:00 200.6 F 0.0
15:00 199.0 F 0.0
20:00 195.5 F 0.0
25:00 192.9 F 0.0
30:00 190.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

04/05/2024 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____