
Midmark M11 - v1.0.0
Total Cycles: 4677

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2024 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 175.6 F 2.3
4:00 177.5 F 5.8
6:00 197.0 F 10.0
8:00 215.3 F 6.8
10:00 225.4 F 8.5
12:00 234.1 F 13.2
14:00 236.0 F 9.9
16:00 236.6 F 8.9
18:00 246.7 F 13.8
20:00 257.9 F 19.8
22:00 267.9 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.6
2:00 272.6 F 29.2
3:00 271.4 F 28.7
4:00 271.2 F 28.9

Min 271.0 F 28.6
Max 273.1 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 226.9 F 0.0
10:00 237.1 F 0.0
15:00 230.7 F 0.0
20:00 230.3 F 0.0
25:00 230.1 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 23:30
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

06/26/2024 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____