
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002153
7:36 17/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:47
136.0 C 220kPa 3:47
136.3 C 224kPa 4:17
136.2 C 222kPa 4:47
136.1 C 222kPa 5:17
136.3 C 223kPa 5:47
136.2 C 221kPa 6:17
136.4 C 225kPa 6:47
136.4 C 224kPa 7:17
136.4 C 224kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 38:08

Digital Signature #
3114B606E4E5AE0F
