

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3783

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2024 02:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 167.0 F 2.6  
4:00 174.6 F 6.4  
6:00 192.5 F 10.8  
8:00 214.5 F 5.2  
10:00 222.6 F 9.4  
12:00 232.0 F 10.5  
14:00 238.4 F 11.7  
16:00 238.1 F 10.4  
18:00 250.8 F 16.7  
20:00 262.0 F 23.1  
22:00 272.3 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.0  
1:00 272.1 F 30.2  
2:00 272.3 F 30.4  
3:00 272.3 F 30.4  
4:00 272.4 F 30.4

Min 271.9 F 29.9  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 217.8 F 0.0  
10:00 218.7 F 0.0  
15:00 217.0 F 0.0  
20:00 218.3 F 0.0  
25:00 216.0 F 0.0  
30:00 211.2 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 23:09  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:02:22  
CYCLE COMPLETE

07/15/2024 03:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_