

-----  
Midmark M11 - v1.0.5  
Total Cycles: 336

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2021 06:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 72.9 F 0.0  
4:00 97.8 F 0.1  
6:00 138.8 F 0.5  
8:00 146.7 F 3.3  
10:00 165.1 F 6.5  
12:00 207.5 F 10.2  
14:00 206.5 F 6.4  
16:00 216.5 F 7.7  
18:00 230.3 F 11.9  
20:00 231.4 F 8.6  
22:00 240.7 F 12.0  
24:00 243.0 F 12.7  
26:00 255.3 F 18.9  
28:00 265.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.2 F 29.8  
2:00 271.8 F 30.0  
3:00 271.7 F 29.4  
4:00 272.3 F 30.1

Min 271.2 F 29.2  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 197.7 F 0.0  
10:00 201.6 F 0.0  
15:00 210.8 F 0.0  
20:00 208.5 F 0.0  
25:00 206.9 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 29:44  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:07:44  
CYCLE COMPLETE

09/08/2021 07:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01