
Midmark M11 - v1.0.5
Total Cycles: 1161

Name: RITTER2
Sterilizer ID: V2313228
06/06/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 184.9 F 3.5
4:00 195.7 F 7.7
6:00 211.5 F 13.0
8:00 223.5 F 7.4
10:00 238.1 F 12.8
12:00 237.1 F 10.4
14:00 235.9 F 9.0
16:00 250.1 F 16.0
18:00 261.9 F 22.8
20:00 272.7 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.2 F 30.1
2:00 272.6 F 29.8
3:00 271.7 F 29.3
4:00 271.2 F 29.0

Min 271.1 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 231.7 F 0.0
10:00 231.3 F 0.0
15:00 229.5 F 0.0
20:00 227.0 F 0.0
25:00 229.3 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

06/06/2022 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____