
Midmark M11 - v1.0.5
Total Cycles: 458

Name: RITTER2
Sterilizer ID: V2313228
10/19/2021 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 73.6 F 0.0
4:00 78.8 F 0.0
6:00 92.4 F 0.1
8:00 135.6 F 0.2
10:00 174.9 F 0.4
12:00 199.2 F 3.5
14:00 209.4 F 6.7
16:00 219.7 F 10.3
18:00 226.5 F 9.4
20:00 234.1 F 10.4
22:00 234.3 F 8.9
24:00 244.6 F 13.1
26:00 244.1 F 12.7
28:00 256.0 F 18.9
30:00 266.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.6 F 29.7
2:00 271.5 F 28.8
3:00 272.7 F 29.4
4:00 271.6 F 29.0

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 190.0 F 0.0
10:00 218.5 F 0.0
15:00 228.2 F 0.0
20:00 219.0 F 0.0
25:00 204.7 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:32
CYCLE COMPLETE

10/19/2021 09:12 AM
ACCEPTED: X
REJECTED:
Operator: U01