
Midmark M11 - v1.0.5
Total Cycles: 1341

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 188.1 F 1.0
4:00 201.5 F 6.2
6:00 218.5 F 11.4
8:00 227.8 F 7.7
10:00 239.8 F 13.4
12:00 241.2 F 11.8
14:00 240.8 F 11.1
16:00 254.1 F 17.9
18:00 265.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.6 F 28.8
3:00 272.3 F 29.7
4:00 272.2 F 29.4

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 234.4 F 0.0
10:00 227.6 F 0.0
15:00 229.8 F 0.0
20:00 227.0 F 0.0
25:00 230.7 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

07/25/2022 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____