
Midmark M11 - v1.0.5
Total Cycles: 2891

Name: RITTER2
Sterilizer ID: V2313228
09/15/2023 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.7 F 0.0
2:00 116.7 F 0.0
4:00 141.5 F 0.0
6:00 173.5 F 1.8
8:00 187.9 F 5.0
10:00 203.9 F 8.4
12:00 216.4 F 12.1
14:00 226.0 F 6.4
16:00 231.7 F 10.6
18:00 233.4 F 8.4
20:00 242.6 F 12.4
22:00 241.1 F 11.0
24:00 252.8 F 16.9
26:00 263.1 F 22.9
28:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.1 F 29.0
2:00 271.5 F 29.0
3:00 271.7 F 28.8
4:00 271.7 F 29.2

Min 271.0 F 28.4
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 228.1 F 0.0
10:00 191.5 F 0.0
15:00 186.7 F 0.0
20:00 181.9 F 0.0
25:00 178.5 F 0.0
30:00 176.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:23
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:06:58
CYCLE COMPLETE

09/15/2023 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____