
Midmark M11 - v1.0.5
Total Cycles: 4431

Name: RITTER2
Sterilizer ID: V2313228
03/17/2025 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 73.1 F 0.2
4:00 93.6 F 0.8
6:00 102.4 F 3.0
8:00 144.7 F 6.1
10:00 171.0 F 9.5
12:00 186.8 F 13.2
14:00 199.7 F 4.8
16:00 212.5 F 8.1
18:00 225.1 F 11.9
20:00 227.2 F 7.1
22:00 238.8 F 12.0
24:00 236.6 F 9.5
26:00 248.5 F 15.3
28:00 259.1 F 21.2
30:00 268.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 272.2 F 29.8
2:00 272.6 F 29.6
3:00 271.9 F 29.5
4:00 272.8 F 29.9

Min 271.5 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 184.3 F 0.0
10:00 194.4 F 0.0
15:00 189.7 F 0.0
20:00 188.7 F 0.0
25:00 185.0 F 0.0
30:00 181.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:04
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

03/17/2025 09:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____